

**CORRIGENDUM No. 1  
to the**

**TENDER DOSSIER**

**Reference: PHASE\_2023-TD1**

**Subject:** Procurement of medical and IT equipment within PHASE project

**Location:** Podgorica, Montenegro

**Reason for change**

Modification of original information submitted by the contracting authority.

**Text to be corrected in the original tender dossier**

Annex II + III: Technical Specifications + Technical Offer

Lot No. 1: Medical equipment

Instead of:

Lot No. 1, Item Number 1.1

<b>1. Item number</b>	<b>2. Specifications required</b>
<b>1.1</b>	<b>Electric hospital bed with 4 motors</b> <b>Quantity: 4</b> Linear motors with min. IPX4 protection for quiet operation <ul style="list-style-type: none"><li>• Electrically adjustable back, legs, height, trendelenburg and counter trendelenburg</li><li>• Dimensions: 103.5cm*212.5cm</li><li>• Height adjustment: 40-73cm</li><li>• Adjustment of the lying section: 70°</li><li>• Leg adjustment: 30°</li><li>• Trendelenburg / Anti-Trendelenburg: 12°</li><li>• Load capacity: 230 kg</li><li>• Remote control</li><li>• Nurse panel with locking function</li><li>• Manual and electronic CPR</li><li>• Removable polypropylene (PP) head and foot panel</li><li>• 4-part lying surface made of ABS material</li><li>• Battery</li><li>• 360° rotating wheels with central locking (diameter 150mm)</li><li>• Plastic protective bumpers against impacts</li><li>• Infusion stand adjustable in height</li><li>• Electrostatic painted metal parts</li><li>• Angle indicator on the sides</li><li>• Mattress 85x190x12 cm 28DNS</li></ul>

Read:

Lot No. 1, Item Number 1.1

1. Item number	2. Specifications required
1.1	<p><b>Electric hospital bed with 4 motors</b>  <b>Quantity: 4</b>            Linear motors with min. IPX4 protection for quiet operation</p> <ul style="list-style-type: none"> <li>• Electrically adjustable back, legs, height, trendelenburg and counter trendelenburg</li> <li>• Dimensions: 103.5cm*212.5cm</li> <li>• Height adjustment: 45-75cm (45 cm or below - up to 75 cm)</li> <li>• Adjustment of the lying section: 70°</li> <li>• Leg adjustment: 30°</li> <li>• Trendelenburg / Anti-Trendelenburg: 12°</li> <li>• Load capacity: 230 kg</li> <li>• Remote control</li> <li>• Nurse panel with locking function</li> <li>• Manual and electronic CPR</li> <li>• Removable polypropylene (PP) head and foot panel</li> <li>• 4-part lying surface made of ABS material</li> <li>• Battery</li> <li>• 360° rotating wheels with central locking (diameter 150mm)</li> <li>• Plastic protective bumpers against impacts</li> <li>• Infusion stand adjustable in height</li> <li>• Electrostatic painted metal parts</li> <li>• Angle indicator on the sides</li> <li>• Mattress 85x190x12 cm 28DNS</li> </ul>

Instead of:

Lot No. 1, Item Number 1.2

1. Item number	2. Specifications required
1.2	<p><b>Electric hospital bed with 3 motors</b>  <b>Quantity: 41</b></p> <ul style="list-style-type: none"> <li>- Bed dimensions: length 214.5cm, width 108cm</li> <li>- Raising the bed in the range of 41-76 cm</li> <li>- Electrical adjustment of the back part from 0 to 70°</li> <li>- Electric adjustment of the foot part from 0 - 30°</li> <li>- Steel construction coated with an electrostatic and antibacterial layer</li> <li>- 4 sections under the mattress covered with an ABS plastic layer</li> <li>- The bed can support a patient's weight of 200 kg</li> <li>- It has 4 protective plastic wheel-bumpers on the corners</li> <li>- It has an anti-decubitus mattress dimensions 190 x 90 x 11 cm</li> <li>- Mattress density 30 kg/m<sup>3</sup></li> <li>- Urine bag holders on the sides of the bed with hooks</li> <li>- 125mm diameter wheel with 360° rotation</li> <li>- Plastic fences on all sides of the bed</li> <li>- Infusion stand, with 4 plastic hooks adjustable in height</li> <li>- Handle for CRP</li> <li>- Maximum load capacity 200kg</li> <li>- Length extension in the foot part of the bed 20 cm</li> <li>- Bed weight 115 kg</li> </ul>

Read:

Lot No. 1, Item Number 1.2

1. Item number	2. Specifications required
1.2	<b>Electric hospital bed with 3 motors</b> <b>Quantity: 41</b> <ul style="list-style-type: none"><li>- Bed dimensions: length 214.5cm, width 108cm</li><li>- Raising the bed in the range of 50-76 cm (50 cm or below - up to 76 cm)</li><li>- Electrical adjustment of the back part from 0 to 70°</li><li>- Electric adjustment of the foot part from 0 - 30°</li><li>- Steel construction coated with an electrostatic and antibacterial layer</li><li>- 4 sections under the mattress covered with an ABS plastic layer</li><li>- The bed can support a patient's weight of 200 kg</li><li>- It has 4 protective plastic wheel-bumpers on the corners</li><li>- It has an anti-decubitus mattress dimensions 190 x 90 x 11 cm</li><li>- Mattress density 28 kg/m<sup>3</sup></li><li>- Urine bag holders on the sides of the bed with hooks</li><li>- 125mm diameter wheel with 360° rotation</li><li>- Plastic fences on all sides of the bed</li><li>- Infusion stand, with 4 plastic hooks adjustable in height</li><li>- Handle for CRP</li><li>- Maximum load capacity 200kg</li><li>- Length extension in the foot part of the bed 20 cm</li></ul>

Instead of:

Lot No. 1, Item Number 1.5

1. Item number	2. Specifications required
1.5	<b>Pulse Oximeter</b> <b>Quantity: 2</b> <ul style="list-style-type: none"><li>-Display: SpO<sub>2</sub>, Pulse Rate;</li><li>-Oxygen Saturation Measuring Range: 35 – 100%;</li><li>-Heart Rate Measuring Range: 30 – 240 BPM;</li><li>-Display: Single-Colour LED;</li><li>-Automatic Shut down after 5 sec when not in use;</li></ul>

Read:

Lot No. 1, Item Number 1.5

1. Item number	2. Specifications required
1.5	<b>Pulse Oximeter</b> <b>Quantity: 2</b> <ul style="list-style-type: none"><li>-Display: SpO<sub>2</sub>, Pulse Rate;</li></ul>

	<ul style="list-style-type: none"><li>-Oxygen Saturation Measuring Range: 45 – 100%;</li><li>-Heart Rate Measuring Range: 30 – 240 BPM;</li><li>-Display: Single-Colour LED;</li><li>-Automatic Shut down after 5 sec when not in use;</li></ul>
--	--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------

All other terms and conditions of the Technical Specifications and the entire Tender Dossier remain unchanged. The above modifications to the Technical Specifications are integral part of the Technical Specifications and the Tender Dossier.